



Acq. Operator : SYSTEM

Seq. Line :

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P2-B-02

Injection Date : 1/13/2022 8:06:33 AM

Inj : 1

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\01132201\01132201 2022-01-13  
05-53-12\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax  
Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity  
II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

