



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :   11
Acq. Instrument : LCMS-1                             Location  :   P1-B2
Injection Date  : 6/5/2019 4:41:11 PM                Inj       :    1
                                                    Inj Volume: 1.000 µl

Acq. Method     : D:\DATA\06051901\06051901 2019-06-05 13-34-24\Purity195_315.M
Last changed    : 6/5/2019 1:34:24 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed    : 4/10/2019 10:56:50 AM by SYSTEM

Method Info     : Operator: GA
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                 /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                 Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                 Reference.
  
```

Sample Info : MCPA acid

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

