

File :Q:\GCMS-06 Minimal\DATA\121620\28- KM.D  
Operator : Organic QC Lab  
Acquired : 16 Dec 2020 19:10 using AcqMethod CHICK\_2019.M  
Instrument : GCMS 6  
Sample Name: P-172S  
Misc Info : O,O,O-Triethylphosphorothioate in MeOH  
Vial Number: 48



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

