

File :Q:\GCMS-08 Minimal\DATA\20191126\24-.D
Operator : KM / JP / ZA / YX
Acquired : 26 Nov 2019 17:16 using AcqMethod KERLA-2017_S100.M
Instrument : GCMS 8
Sample Name: P-208S-H-10X
Misc Info : Terbufos in Hexane
Vial Number: 134



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

