

Print of window 38: Current Chromatogram(s)

Data File : D:\DATA\06141801\06141801 2018-06-14 13-11-30\04- Sample Name :  
P-216S-CN-10X (Carboxin)



=====

|                 |                        |            |            |
|-----------------|------------------------|------------|------------|
| Acq. Operator   | : SYSTEM               | Seq. Line  | : 4        |
| Acq. Instrument | : LCMS-1               | Location   | : P1-C1    |
| Injection Date  | : 6/14/2018 2:08:06 PM | Inj        | : 1        |
|                 |                        | Inj Volume | : 1.000 µl |

Acq. Method : D:\DATA\06141801\06141801 2018-06-14 13-11-30\Purity195\_315.M  
Last changed : 6/14/2018 1:11:30 PM by SYSTEM  
Analysis Method : C:\Chem32\1\Methods\Purity195\_315.M  
Last changed : 6/6/2017 9:45:25 AM by SYSTEM  
Method Info : Operator: GA  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water /100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min . Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

