

File :D:\MassHunter\GCMS\1\DATA\03022101\06- KD.D
Operator : Organic QC department
Acquired : 02 Mar 21 09:02 am using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: P-226S
Misc Info :
Vial Number: 38



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

