

=====

Acq. Operator	: SYSTEM	Seq. Line	: 1
Sample Operator	: SYSTEM		
Acq. Instrument	: LCMS-2	Location	: P1-A1
Injection Date	: 2/11/2020 2:57:00 PM	Inj	: 1
		Inj Volume	: 2.000 µl
Acq. Method	: C:\Users\Public\Documents\ChemStation\1\Data\02112001\01202001 2020-02-11 14-56-00\Diquat.M		
Last changed	: 2/11/2020 2:54:14 PM by SYSTEM		
Analysis Method	: C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M		
Last changed	: 2/11/2020 3:39:52 PM by SYSTEM (modified after loading)		
Method Info	: General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.		

Current Chromatogram(s)

