Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\03202001\02- .D
Sample Name : P-263S

AccuStandard®
Leader in Analytical Reference Standards

Acq. Operator	: SYSTEM	Seq. Line	: 2
Acq. Instrument	: LCMS-1	Location	: P1-E2
Injection Date	: 3/20/2020 6:28:00 AM	Inj	: 2
		Inj Volume	: 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 10/18/2019 4:18:14 PM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Danitol in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)