



=====

| | | | |
|-----------------|-------------------------|------------|------------|
| Acq. Operator | : SYSTEM | Seq. Line | : 5 |
| Acq. Instrument | : LCMS-1 | Location | : P1-F6 |
| Injection Date | : 7/11/2018 12:24:45 PM | Inj | : 1 |
| | | Inj Volume | : 1.000 µl |

Acq. Method : D:\DATA\07101801\07101801 2018-07-11 11-09-15\Purity195_315.M
Last changed : 7/11/2018 11:09:15 AM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 7/11/2018 9:56:54 AM by SYSTEM
(modified after loading)

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

