

=====

Acq. Operator : SYSTEM	Seq. Line : 23
Acq. Instrument : LCMS-1	Location : P2-A7
Injection Date : 2/24/2022 4:30:33 PM	Inj : 1
	Inj Volume : 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 11/5/2021 2:37:17 PM by SYSTEM  
Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Sample Info : Mecarbam @ 100ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

