

File : L:\QC Data\GCMS-08 Minimal\DATA\20220523\34 -SAK.D
Operator : Organic QC Lab
Acquired : 23 May 2022 19:16 using AcqMethod Hy_Pest_24cm_S10.M
Instrument : GCMS 8
Sample Name : P-328S
Misc Info : resample
Vial Number : 34



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

