

File :Q:\GCMS-06 Minimal\DATA\020821\21- -DC.D
Operator : Organic QC Lab
Acquired : 08 Feb 21 08:12 pm using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: P-331S-H
Misc Info : Oxychlordanes Isomer in Hexane
Vial Number: 28



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

