



=====  
 Acq. Operator : SYSTEM Seq. Line : 9  
 Sample Operator : SYSTEM  
 Acq. Instrument : LCMS-2 Location : P1-A-08  
 Injection Date : 6/30/2020 6:38:22 PM Inj : 1  
 Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06302001\06302001 2020-06-30 16-05-59  
 \Purity\_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

