

File :Q:\GCMS-07 Minimal\DATA\06302001\11- KM.D  
Operator : Organic QC department  
Acquired : 30 Jun 2020 16:12 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-379S  
Misc Info : Iodofenphos in MeOH  
Vial Number: 56



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

