



```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    4
Acq. Instrument : LCMS-1                               Location  : P2-j
Injection Date  : 1/5/2021 12:42:05 PM                 Inj       :    1
                                                    Inj Volume: 1.000 µl
```

```
Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 5/8/2020 11:20:44 AM by SYSTEM
Analysis Method: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 11/30/2020 12:15:35 PM by SYSTEM
```

```
(modified after loading)
Method Info    : Operator: Oraganic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                Reference.
```

Additional Info : Peak(s) manually integrated

### Current Chromatogram(s)

