

=====

| | | | |
|-----------------|------------------------|------------|------------|
| Acq. Operator | : SYSTEM | Seq. Line | : 4 |
| Acq. Instrument | : LCMS-1 | Location | : P1-E8 |
| Injection Date | : 1/20/2021 7:56:06 AM | Inj | : 1 |
| | | Inj Volume | : 1.000 µl |

Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/8/2020 11:20:44 AM by SYSTEM
Method Info : Operator: Oraganic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

