



Acq. Operator : SYSTEM Seq. Line : 49
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-E-04
Injection Date : 1/24/2025 11:12:51 PM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\012425\012425 2025-01-24 08-07-07
\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

