



```
=====
Acq. Operator   : SYSTEM                      Seq. Line :
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  : P1-E-01
Injection Date  : 5/20/2022 7:30:41 AM        Inj       : 2
                                                Inj Volume: 1.000 µl
```

```
Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\05202201\05192201 2022-05-20
                  07-10-48\Purity_195-315.M
Last changed    : 3/23/2022 11:04:54 AM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed    : 5/16/2022 12:35:31 PM by SYSTEM
                  (modified after loading)
```

```
Method Info     : OPERATOR: Organic QC lab
                  General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%
                  Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
                  Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity
                  II LC/MS system, Agilent 6125 quadrupole MS detector.
```

Sample Info : Chlortoluron in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

