

=====

Acq. Operator : SYSTEM	Seq. Line : 21
Acq. Instrument : LCMS-1	Location : P2-A1
Injection Date : 6/21/2023 12:19:24 PM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 1/5/2023 12:01:53 PM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 6/23/2023 10:36:19 AM by SYSTEM  
(modified after loading)

Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml  
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size.  
Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple MS Detector. No UV  
Reference.

Sample Info : 1-Naphthylacetic acid in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

