

=====

Acq. Operator : SYSTEM	Seq. Line : 3
Acq. Instrument : LCMS-1	Location : P1-C2
Injection Date : 11/4/2020 7:26:35 AM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl

Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/8/2020 11:20:44 AM by SYSTEM
Method Info : Operator: Oraganic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

