Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\03252022\13- -KD.D
Sample Name : P-505S-W

Acq. Operator	: SYSTEM	Seq. Line	: 13
Acq. Instrument	: LCMS-1	Location	: P1-F7
Injection Date	: 3/25/2022 1:43:39 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 3/16/2022 2:44:36 PM by SYSTEM
(modified after loading)
Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Sodium diethyldithiocarbamate @ 100ug/mL in Water

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)