

=====

Acq. Operator	: SYSTEM	Seq. Line	: 4
Sample Operator	: SYSTEM		
Acq. Instrument	: LCMS-2	Location	: P2-A-08
Injection Date	: 4/7/2020 11:34:52 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\04072001\04062001 2020-04-07  
10-37-05\Purity\_195-315.M

Last changed : 12/3/2019 1:25:24 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 4/7/2020 4:22:47 PM by SYSTEM  
(modified after loading)

Method Info : General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax  
Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity  
II LC/MS system, Agilent 6125 quadrupole MS detector.



Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

