



=====  
Acq. Operator : SYSTEM Seq. Line : 19  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2 Location : P2-F1  
Injection Date : 10/25/2021 3:15:55 PM Inj : 1  
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\10252101\10252101 2021-10-25  
09-35-01\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax  
Eclipse plus C-18, 2.1mmx50mm, 1.8 µ particle size. Agilent 1290 Infinity  
II LC/MS system, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

