



```

=====
Acq. Operator   : SYSTEM                      Seq. Line   :   35
Acq. Instrument : LCMS-1                      Location    :   P1-B2
Injection Date  : 6/20/2018 2:25:12 AM      Inj         :    1
                                                    Inj Volume  : 1.000 µl
Acq. Method     : D:\DATA\06191801\06191801 2018-06-19 15-51-20\Purity195_315.M
Last changed    : 6/19/2018 3:51:20 PM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed    : 6/18/2018 9:35:46 AM by SYSTEM
                (modified after loading)
Method Info     : Operator: GA
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.
                3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule
                size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector.
                No UV Reference.
    
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

