

```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    4
Acq. Instrument : LCMS-1                     Location  : P1-C4
Injection Date  : 7/14/2022 7:52:45 AM      Inj       :    1
                                           Inj Volume: 1.000 µl

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 7/12/2022 8:39:31 AM by SYSTEM
                 (modified after loading)

Method Info     : Operator: Organic QC lab
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                 :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                 particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                 MS Detector. No UV Reference.
=====
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

