



=====

Acq. Operator : SYSTEM	Seq. Line : 12
Acq. Instrument : LCMS-1	Location : P1-D5
Injection Date : 5/25/2016 2:25:09 PM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 10.000 µl  
Acq. Method : D:\DATA\05251601\05231601 2016-05-25 10-59-12\POS\_NEG\_50\_50\_GRAD0-100\_ NoRef.M

Last changed : 5/25/2016 10:59:12 AM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\POS\_NEG\_50\_50\_GRAD0-100.M  
Last changed : 10/30/2017 12:22:23 PM by SYSTEM  
(modified after loading)

Method Info : Operator: GA  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water /100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min . Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

