

=====

Acq. Operator : SYSTEM Seq. Line : 24

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P2-C-01

Injection Date : 4/17/2024 7:45:35 PM Inj : 1

Inj Volume : 1.000 µl



Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\041724\041724 2024-04-17 12-31-07\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µ particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : 3,5,6-Trichloro-2-pyridinol @ 100 ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

