



```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  : P2-A-01
Injection Date  : 12/18/2023 7:30:32 AM              Inj       :    1
                                                    Inj Volume: 1.000 µl
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

```
Acq. Method      : C:\Users\Public\Documents\ChemStation\1\Data\121823\121823 2023-12-18 07-10
                  -48\Purity_195-315.M
Last changed     : 2/6/2023 2:02:18 PM by SYSTEM
Analysis Method  : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed     : 2/6/2023 2:02:18 PM by SYSTEM
Method Info      : OPERATOR: Organic QC lab
```

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

