

=====

Acq. Operator	: SYSTEM	Seq. Line	: 5
Acq. Instrument	: LCMS-1	Location	: P2-E9
Injection Date	: 6/17/2022 8:29:35 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
 Last changed : 11/5/2021 2:37:17 PM by SYSTEM  
 Method Info : Operator: Organic QC lab  
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate  
 :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron  
 particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple  
 MS Detector. No UV Reference.

Sample Info : Ethirimol @ 100ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

