

File :D:\MassHunter\GCMS\1\data\20180626\33-
Operator : JJP / KM
Acquired : 27 Jun 2018 01:37 using AcqMethod Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: P-646S
Misc Info : Barnon in MeOH
Vial Number: 31



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

