

File :Q:\GCMS-06 Minimal\DATA\040820\33-.D  
Operator : Organic QC Lab  
Acquired : 08 Apr 2020 21:11 using AcqMethod CHICK\_2019\_2S\_2uL.M  
Instrument : GCMS 6  
Sample Name: P-653S-TP-0.1X  
Misc Info : Ofurace in Isooctane  
Vial Number: 56



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

