

File :D:\MassHunter\GCMS\1\data\20180627\05-
Operator : JJP / KM
Acquired : 27 Jun 2018 21:36 using AcqMethod Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: P-661S-A-10X
Misc Info : Isoprothiolane 1mg/ml in acetone
Vial Number: 5



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

