

```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Acq. Instrument : LCMS-1                     Location  : P1-F4
Injection Date  : 3/18/2021 10:40:45 AM      Inj       :    1
                                                    Inj Volume: 1.000 µl

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 5/8/2020 11:20:44 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 3/12/2021 4:31:42 PM by SYSTEM
                (modified after loading)

Method Info     : Operator: Oraganic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                MS Detector. No UV Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

