

=====

Acq. Operator	: SYSTEM	Seq. Line	: 4
Acq. Instrument	: LCMS-1	Location	: P1-A9
Injection Date	: 7/13/2018 4:42:47 PM	Inj	: 1

Inj Volume : 1.000 µl  
Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl  
Acq. Method : D:\DATA\07131801\07101801 2018-07-13 15-46-15\Purity195\_315.M  
Last changed : 7/13/2018 3:46:15 PM by SYSTEM  
Analysis Method : C:\Chem32\1\Methods\Purity195\_315.M  
Last changed : 6/6/2017 9:45:25 AM by SYSTEM  
Method Info : Operator: GA  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

