



Acq. Operator : SYSTEM  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2  
Injection Date : 9/21/2021 2:47:48 PM  
Seq. Line :  
Location : P1-C-05  
Inj : 1  
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\09212101\09212101 2021-09-21 06-16-08  
\Purity\_195-315.M  
Last changed : 4/29/2021 10:41:07 AM by SYSTEM  
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M  
Last changed : 4/29/2021 10:41:07 AM by SYSTEM  
Method Info : OPERATOR: Organic QC lab  
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system  
, Agilent 6125 quadrupole MS detector.

Sample Info : Tribenuron-methyl @ 100ug/mL in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

