

File :D:\MassHunter\GCMS\1\data\05221901\08- D
Operator : KD
Acquired : 22 May 19 08:07 am using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: P-683S-CN
Misc Info : Chlozolate in ACCN
Vial Number: 48

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS: INJ 250c.

