Print of window 38: Current Chromatogram(s)
Data File : C:\USERS\P...ON\1\DATA\06202201\06202201 2022-06-20 10-43-36\07- 222061355 ALR.D
Sample Name : P-692S

==========

Acq. Operator : SYSTEM Seq. Line :
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P1-A-08
Injection Date : 6/20/2022 12:37:45 PM Inj : 1
Inj Volume : 1.000 µl
Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06202201\06202201 2022-06-20 10-43-36\Purity_195-315.M
Last changed : 3/23/2022 11:04:54 AM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed : 3/23/2022 11:04:54 AM by SYSTEM
Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min,the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Quinclorac @ 100ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)