

=====

Acq. Operator	: SYSTEM	Seq. Line	: 31
Acq. Instrument	: LCMS-1	Location	: P1-A7
Injection Date	: 6/20/2018 1:10:47 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : D:\DATA\06191801\06191801 2018-06-19 15-51-20\Purity195\_315.M  
Last changed : 6/19/2018 3:51:20 PM by SYSTEM  
Analysis Method : C:\Chem32\1\Methods\Purity195\_315.M  
Last changed : 6/18/2018 9:35:46 AM by SYSTEM  
(modified after loading)

Method Info : Operator: GA  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

