

=====

Acq. Operator : SYSTEM	Seq. Line : 26
Acq. Instrument : LCMS-1	Location : P1-F4
Injection Date : 12/15/2022 6:43:00 PM	Inj : 1
	Inj Volume : 1.000 µl
Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M	
Last changed : 11/5/2021 2:37:17 PM by SYSTEM	
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M	
Last changed : 12/16/2022 9:39:19 AM by SYSTEM	
	(modified after loading)
Method Info : Operator: Organic QC lab	
	Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
	Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
	:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
	particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
	MS Detector. No UV Reference.



Sample Info : Sulfentrazone in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

