

=====

|                                       |                       |
|---------------------------------------|-----------------------|
| Acq. Operator : SYSTEM                | Seq. Line : 8         |
| Acq. Instrument : LCMS-1              | Location : P1-B8      |
| Injection Date : 7/28/2020 7:11:33 AM | Inj : 1               |
|                                       | Inj Volume : 1.000 µl |

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 5/8/2020 11:20:44 AM by SYSTEM  
Method Info : Operator: Oraganic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate  
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron  
particulate size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole  
MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

