



```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    5
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                      Location  : P2-B-07
Injection Date  : 12/6/2021 8:56:51 AM        Inj       :    1
                                           Inj Volume: 1.000 µl
```

```
Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\12062101\12062101 2021-12-06 07-40-28
                  \Purity_195-315.M
```

```
Last changed    : 4/29/2021 10:41:07 AM by SYSTEM
```

```
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
```

```
Last changed    : 12/1/2021 8:27:34 AM by SYSTEM
                  (modified after loading)
```

```
Method Info     : OPERATOR: Organic QC lab
```

```
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.
```

```
Additional Info : Peak(s) manually integrated
```

Current Chromatogram(s)

