

=====

Acq. Operator	: SYSTEM	Seq. Line	: 14
Acq. Instrument	: LCMS-1	Location	: P1-A9
Injection Date	: 6/30/2022 3:49:50 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 6/21/2022 4:09:45 PM by SYSTEM
(modified after loading)
Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
MS Detector. No UV Reference.

Sample Info : Picoxystrobinin AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

