

Acq. Operator : SYSTEM Seq. Line : 10
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P1-A-09
Injection Date : 4/15/2020 8:50:27 AM Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\04152001\04062001 2020-04-15 05-59-11
\Purity_195-315.M

Last changed : 12/3/2019 1:25:24 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 4/10/2020 12:43:38 PM by SYSTEM
(modified after loading)

Method Info : General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µ particlue size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Sample Info : Flonicamid in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

