



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :    6
Sample Operator : SYSTEM                               Location  : P1-F-03
Acq. Instrument : LCMS-2                              Inj       :    1
Injection Date  : 10/21/2020 12:56:59 PM            Inj Volume: 1.000 µl
    
```

```

Acq. Method    : C:\Users\Public\Documents\ChemStation\1\Data\10202001\10152001 2020-10-21 11-21-26
                 \Purity_195-315.M
    
```

```

Last changed   : 5/8/2020 10:12:19 AM by SYSTEM
    
```

```

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
    
```

```

Last changed   : 10/20/2020 2:10:33 PM by SYSTEM
                 (modified after loading)
    
```

```

Method Info    : OPERATOR: Organic QC lab
    
```

```

General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%
Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.
    
```

```

Additional Info : Peak(s) manually integrated
    
```

Current Chromatogram(s)

