

=====

Acq. Operator	: SYSTEM	Seq. Line	: 31
Acq. Instrument	: LCMS-1	Location	: P1-A8
Injection Date	: 2/26/2024 6:44:06 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 1/5/2023 12:01:53 PM by SYSTEM  
Method Info : Operator: Organic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple MS Detector. No UV Reference.

Sample Info : Cyclanilide in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

