

File :Q:\GCMS-07\DATA\06121801\10- : K.D
Operator
Acquired : 12 Jun 2018 11:17 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: P-997S-A-10X
Misc Info : Fenoxanil @1000ug/mL in Acetone
Vial Number: 9

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

