

File :D:\MassHunter\GCMS\1\data\20230227\17- =JP=.D  
Operator : Organic QC Lab  
Acquired : 28 Feb 23 12:07 pm using AcqMethod Hy\_Pest\_24cm\_S2\_2u1inj.M  
Instrument : GCMS 8  
Sample Name: PC-001S  
Misc Info : Octachlorostyrene; in toluene  
Vial Number: 6



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

