

File :D:\MassHunter\GCMS\1\data\20190130\30- D
Operator : JJP / KM
Acquired : 31 Jan 19 09:30 am using AcqMethod KERLA-2Split-2uL.M
Instrument : GCMS 8
Sample Name: PCB-DUTCH7
Misc Info : Dutch Seven PCB Standard in Isooctane
Vial Number: 70



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 2:1, 2uL inj.; GC/MS; INJ 270c

