

File :M:\GCMS-07 Minimal\DATA\09232101\37- ALR.D  
Operator : Organic QC department  
Acquired : 24 Sep 2021 05:03 using AcqMethod KERLA-2Split-2uL.M  
Instrument : 5977B MSD  
Sample Name: PCB-SIM  
Misc Info : PCB/SIM Solution in Hexane  
Vial Number: 119



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

