



```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Sample Operator : SYSTEM                      Location  :   P2-E-06
Acq. Instrument : LCMS-2                      Inj       :    2
Injection Date  : 6/4/2024 7:38:52 AM        Inj Volume: 1.000 µl
=====
```

```
Acq. Method    : C:\Users\Public\Documents\ChemStation\1\Data\060424\060424 2024-06-04 07-19-07\Purity_195-315.M
Last changed   : 2/6/2023 2:02:18 PM by SYSTEM
Analysis Method: C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed   : 5/30/2024 7:55:21 AM by SYSTEM (modified after loading)
```

```
Method Info    : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

